

iFlash-SARS-CoV-2 IgG reference control value

| Product | Analyte | Lot No. | Specification | Expiration date | Ref. value | SD | Ref. range | Unit |
|------------------------------|----------------|----------|---------------|-----------------|------------|------|-------------|-------|
| iFlash-SARS-CoV-2 IgG Kit QC | SARS-CoV-2 IgG | 20200201 | Negative | 2024/02/19 | 5.51 | / | <10 | AU/mL |
| | | | Positive | 2024/02/19 | 19.39 | 2.59 | 11.63-27.15 | AU/mL |

Deshu Cheng
Production Department
2020/06/22

