

**iFlash-SARS-CoV-2 IgG reference control value**

Product	Analyte	Lot No.	Specification	Expiration date	Ref. value	SD	Ref. range	Unit
iFlash-SARS-CoV-2 IgG Kit QC	SARS-CoV-2 IgG	20200401	Negative	2024/5/6	4.18	/	<10	AU/mL
			Positive	2024/5/6	17.44	2.33	10.46-24.42	AU/mL

Deshu Cheng  
Production Department  
2020/06/23

