

iFlash- 2019-nCoV NAb reference control value

Product Name	REF	Lot	Specification	Expiration date	Value	1SD	Range	Unit
2019-nCoV NAb Control (CLIA)	CC345	20211201	Negative	2023/6/24	5.76	0.76	< 10	AU/mL
			Positive	2023/6/24	38.03	3.54	27.42-48.64	AU/mL

- Product Code No. YH-CC345
- Reagent lot used for the value assignment : 20210901

製造元

YHLO

SHENZHEN YHLO BIOTECH CO., LTD.

Headquarter: Building 1, YHLO Biopark, Baolong 2nd Road, Baolong Subdistrict, Longgang District, 518116 Shenzhen, P.R.China

Tel: +86 755 26601910
Email: marketing@szyhlo.com
Website: www.szyhlo.com

販売元

MBL 株式会社 医学生物学研究所

A JSR Life Sciences Company <https://ruo.mbl.co.jp/>

<お問い合わせ先>

〒105-0012

東京都港区芝大門2丁目11番8号 住友不動産芝大門二丁目ビル

TEL : 03-6684-6860 (代表) FAX : 03-6854-3615

E-mail : kensa@mbl.co.jp